



Agriculture & Horticulture
DEVELOPMENT BOARD



Grower Summary

FV 401

Vining peas: Development of an improved standardisation procedure for the pea tenderometer

Final 2013

Disclaimer

AHDB, operating through its HDC division seeks to ensure that the information contained within this document is accurate at the time of printing. No warranty is given in respect thereof and, to the maximum extent permitted by law the Agriculture and Horticulture Development Board accepts no liability for loss, damage or injury howsoever caused (including that caused by negligence) or suffered directly or indirectly in relation to information and opinions contained in or omitted from this document.

No part of this publication may be reproduced in any material form (including by photocopy or storage in any medium by electronic means) or any copy or adaptation stored, published or distributed (by physical, electronic or other means) without the prior permission in writing of the Agriculture and Horticulture Development Board, other than by reproduction in an unmodified form for the sole purpose of use as an information resource when the Agriculture and Horticulture Development Board or HDC is clearly acknowledged as the source, or in accordance with the provisions of the Copyright, Designs and Patents Act 1988. All rights reserved.

AHDB (logo) is a registered trademark of the Agriculture and Horticulture Development Board. HDC is a registered trademark of the Agriculture and Horticulture Development Board, for use by its HDC division. All other trademarks, logos and brand names contained in this publication are the trademarks of their respective holders. No rights are granted without the prior written permission of the relevant owners.

The results and conclusions in this report may be based on an investigation conducted over one year. Therefore, care must be taken with the interpretation of the results.

Use of pesticides

Only officially approved pesticides may be used in the UK. Approvals are normally granted only in relation to individual products and for specified uses. It is an offence to use non-approved products or to use approved products in a manner that does not comply with the statutory conditions of use, except where the crop or situation is the subject of an off-label extension of use.

Before using all pesticides check the approval status and conditions of use.

Read the label before use: use pesticides safely.

Further information

If you would like a copy of the full report, please email the HDC office (hdc@hdc.ahdb.org.uk), quoting your HDC number, alternatively contact the HDC at the address below.

HDC
Stoneleigh Park
Kenilworth
Warwickshire
CV8 2TL

Tel – 0247 669 2051

HDC is a division of the Agriculture and Horticulture Development Board.

Project Number: FV 401

Project Title: Vining peas: Development of an improved standardisation procedure for the pea tenderometer

Project Leader: Steve Belcher

Contractor: PGRO – Processors and Growers Research Organisation

Industry Representative: Richard Fitzpatrick

Report: Final Report 2013

Publication Date: 09 April 2013

Previous report/(s): -

Start Date: 01 March 2012

End Date: 28 February 2013

Project Cost: £5,349

Headline

The data from this project suggests that the Dodman digital tenderometer can be adopted as the industry standard and be used a reference Master tenderometer against which factory and grower group crosschecks are made.

Background

The Martin pea tenderometer has been the mainstay of the industry as a means of determining pea maturity and payment is usually made according to the tenderometer reading (TR). The tenderometer is used to assess the maturity of crops prior to harvesting and to measure the maturity of the vined produce.

The Martin tenderometer has an upper grid which is moved by means of an electric motor, through a second grid mounted on the same shaft. Peas are filled in the space between the grids and once started, the grids squash the peas and eventually shear through them. The force required is measured by the displacement of a weighted pendulum attached to the second grid and is indicated by a pointer on a scale reading from 0 to 200. There are several points at which the tenderometer can produce variable readings, including the resistance of the weighted pendulum, resistance of worn bearings, slippage between the attachment of the pointer to the grid and movement of the printed scale which is fixed to the back of the grid housing.

Because of these factors, it is necessary to continually adjust the settings on the tenderometer to maintain a standard reading and to avoid drift. PGRO operates a scheme to ensure that Master instruments give accurate results. As a back up, cross checking with the Martin tenderometer at Campden BRI is made several times over the season. Tests should be carried out at the start of the vining season, and then routine tests can be made each week. More frequent tests are made if problems are suspected with a particular tenderometer, or if the factory instrument is used for checking other tenderometers.

The Dodman tenderometer has been introduced as an alternative to the Martin and PGRO undertook some comparison trials in 2003 with the prototype digital instrument which indicated its suitability as a substitute. There were many improvements in the new instrument which resulted in a reduction of the number of moving parts, and a replacement of the mechanical resistance mechanism with one which is directly controlled electronically thereby reducing the variability of the TR values between pea samples.

The most recent digital tenderometer (Dodman) has now been introduced by many processing factories and by larger grower groups. Standardisation is still required for the purposes of quality assurance but it is now necessary for the PGRO service to include the digital tenderometer as part of the standardisation service, and after a full validation, replace the master Martin with the Dodman instrument. It is envisaged that the time taken for the standardisation checking will be greatly reduced and the need for regular adjustment to maintain the standard will be reduced. Because of the reduction of reading drift through the season and the fact that the digital instrument can be re set electronically at the start of each season, there will be no need to carry out cross checking with the Campden Martin tenderometer. Such reductions will result in cost saving by the factory and providing growers with the assurance that the tests are comparable, reproducible and reliable through the harvest season.

The current cross check procedure is carried out on a batch of vined, washed vining peas that have been cooled to 20°C. One half of the batch is retained by the test tenderometer operator and the second half is delivered to PGRO. Eleven sub-samples of the mixed sample are measured using the PGRO Master Martin tenderometer, and the mean value calculated after discarding the first reading. At the same time, the test tenderometer operator measures the values on the retained batch of peas using the same methodology. Comparisons between the PGRO Master and the test tenderometer are made and the test tenderometer is adjusted if necessary to calibrate with the PGRO Master.

Summary

At a range of maturities over the range TR 90-130 samples of peas were harvested, vined and washed and the maturity measured by both the PGRO Master Martin tenderometer and the Dodman digital tenderometer.

Over an approximate 3 week period, 29 June to 19 July the maturity of 109 samples of vining peas was measured using the PGRO Master Martin tenderometer and the Dodman tenderometer. This comprised 26 different varieties over a tenderometer range of 79 to 158.5 units (as measured by the Martin tenderometer). Five samples were ran through each machine and the average, maximum, minimum and range of the tenderometer readings calculated.

On a weekly basis over about an eight week period, standardisation of the PGRO Master tenderometer was carried out. The tenderometer reading from a commercially obtained sample of vined, washed and size graded peas was measured. The size grades were small,

medium and large and these served to give a range of maturity. After cooling, half of the samples were retained by PGRO and half sent to Campden BRI in cooled boxes. The maturity of the samples was also measured using the PGRO Dodman tenderometer and a Dodman tenderometer at Campden BRI.

Eleven sub-samples of the mixed samples were measured using the PGRO Master Martin tenderometer, and the average, maximum, minimum and range of the tenderometer readings calculated after discarding the first reading. At the same time, the Campden BRI Master Martin tenderometer and the PGRO and Campden BRI Dodman tenderometers measured the values using the same methodology.

Previous work by PGRO in 2003 with a Dodman prototype tenderometer indicated a good relationship with the PGRO Mater Martin tenderometer. Several factories and grower groups have now adopted the latest version of the Dodman tenderometer and this project aims to show that a Dodman tenderometer could be used as the standard for the industry.

The data within this project suggests that the Dodman digital tenderometer could be adopted as the industry standard and be used a reference Master tenderometer against which factory and grower group crosschecks are made.

Financial Benefits

There are sometimes issues with the calibration of Martin tenderometers. Accurate and consistent estimation of tenderometer readings at the time of delivery to the factory is important to growers as there are financial penalties for pea loads that fall out of the Grade specification.

For example: if peas grown for A grade are down graded to B grade because of tenderometer inaccuracies the financial penalty can be £25 per tonne of peas. At an average yield of 4.95t/ha this equates to a loss of £124/ha.

Action Points

- PGRO to action the use of the Dodman digital tenderometer as the Master for the industry Cross check and standardization.
- Tenderometers need regular maintenance

